

अधीक्षक का कार्यालय

जवाहरलाल नेहरू चिकित्सा महाविद्यालय अस्पताल, भागलपुर।

अल्प कालिन कोटेशन आमंत्रण सूचना ।

Covid-19 & Black Fungus महामारी हेतु जवाहरलाल नेहरू चिकित्सा महाविद्यालय अस्पताल, भागलपुर में ईलाज हेतु HIGH SPEED MICRO DEBRIDER & HAND PIECE के अत्यन्त अति आवश्यकता के क्रम में निबधित आपूर्तिकर्ता/ख्याति प्राप्त फर्मों/कम्पनियों से मुहरबंद कोटेशन आमंत्रित की जाती हैं।

HIGH SPEED MICRO DEBRIDER CONSOLE

SPECIFICATION OF MAIN CONSOLE FOR MICRO DEBRIDER:

1. Should be a versatile powered ENT system, that lets to choose just the power required for various ENT and Aesthetic related surgeries.
2. The system should be suitable for wide variety of procedures ranging from Rhinology, Other transnasal procedures, Otolaryngology / Neuro otology / Nasopharyngeal / laryngeal / tracheal / bronchial.
3. Should drive the hand piece for Adenoid specific application.
4. Should have in built user friendly interactive menu and illustrative help guide.
5. Should have large Touch screen monitor.
6. The various parameters should be able to adjust either from touch screen panel or from the multifunction foot switch.
7. Should be able to connect multiple hand pieces at a time like debrider hand pieces (Up to 7500 RPM in Oscillating mode & 30000 RPM in Forward mode), low speed Otolaryngology drills (Up to 16000 RPM) and High speed Otolaryngology drills (Up to 60000 RPM).
8. Console should recognize the various hand pieces and automatically adjust the settings accordingly.
9. Should have inbuilt pumps each for Irrigation (5Cc / Min to 100Cc / Min) and Cooling.
10. Should have multifunction ergonomically designed foot control with light emission for easy identification.
11. Should be able to control Speed / Mode, Forward / Reverse toggle, Active hand piece change from the Foot control itself.
12. Should have option for remote control Irrigation to operate from sterile area.
13. Should have in built Lens cleaning system.
14. Should have 0 Degree, 30 Degree, 45 Degree & 70 Degree endoscope sheath compatibility for Automatic Lens cleaning system.
15. Main Console should have an option to operate manual foot paddle from console itself in case the foot paddle are not working.
16. Should have the provision to mount the console on various sizes of IV pole.

SPECIFICATION OF HIGH SPEED MICRO DEBRIDER HAND PIECE :

1. Should be compatible with main console.
2. Should be able work up to the speed of 30000 RPM in forward rotation and 7500RPM in oscillation mode.
3. Should have fingertip control to rotate only the tip of the blade up to 360 deg.
4. Should have straight suction path to reduce clogging and allow efficient tissue removal.
5. Should have integrated blade locking system to lock the blade tip rotation.
6. Should have integrated side grooves and cable clips to provide better tubing management.
7. Should have Titanium body to avoid rusting.
8. Should be light in weight and ergonomically designed.
9. Should be able to operate 120 Degree curved blade for Maxillary Sinus Application.
10. Should be able to operate Diamond & Cutting burs for DCR Surgeries.
11. Should have provision to operate Inferior Turbinate blade for Turbinoplasty and Adenoid blade for Adenoidectomy.
12. Should be compatible with Frontal Trephination System for some specific Frontal Sinus Surgery.
Should be compatible to operate Endoscopic Skull Base Bur at 30,000 rpm speed.

ADDED TENDER TERMS

1. The instruments quoted should be of high quality and standard.
2. The Instruments should be imported and of US FDA, CE & ISO certification.
3. Copy of the US FDA Certificate, CE certificate and ISO certificate must be enclosed.
4. All equipment should be manufactured by a single manufacturer & Single vendor shall supply all equipment.
5. At least 2 orders of supply to government organizations should be submitted.

1. अधोहताक्षरी या विभागीय सक्षम पदाधिकारी को बिना कारण बताए कोटेशन रद्द करने का अधिकार सुरक्षित रहेगा।
2. कोटेशन प्राप्ति खोलने की तिथि के दिन किसी कारणवश सरकारी अवकाश घोषित हो जाने पर यह तिथि अगले कार्य दिवस को स्वतः निश्चित समझा जायेगा।
3. कोटेशन दाताओं को मो0 1000 बैंक डी0डी0 रोगी कल्याण समिति, ज.ला.ने.चि.महा.अस्पताल भागलपुर के पद नाम भुगतान कर निविदा कागजात प्राप्त किया जा सकता है।
4. दर GST/IGST के बिना अंकित किया जाये।
5. तकनीकी निविदा के साथ अग्रिम राशि के रूपमें मो0 1,00,000/—(एक लाख) रुपये मात्र का बैंक ड्राफ्ट के साथ रोगी कल्याण समिति, ज.ला.ने.चि.महा.अस्पताल भागलपुर के नाम से देना होगा।
6. निविदाता का एक वर्ष में न्यूनतम पचिस लाख का टर्न ओवर एवं पूर्व तीन वर्ष में 20 लाख का एकल कार्या देश पर आपूर्ति किया हो।
7. निविदा दो प्रतियों में तकनीकी एवं वित्तीय निविदा अलग-अलग मुहरबंद लिफाफे पर कमश तकनीकी एवं वित्तीय निविदा लिखा होना आवश्यक होगा। आपूर्ति किये जाने वाले समायी आदि के कम्पनी द्वारा Authorization Letter तकनीकी निविदा के साथ संलग्न करना अनिवार्य होगा।
8. कोटेशन के साथ अद्दतन आयकर, जी0 एस0 टी0 एवं आयात निर्यात प्रमाणपत्र देना होगा।
9. आपूर्तिकर्ताओं द्वारा गारंटी/ वारंटी उद्घत का दर अंकित करना आवश्यक होगा।
10. विस्तृत जानकारी www.jlnmch.org पर देखें।
11. कोटेशन दाता को आदेशित समायी विभाग में आदेश निगर्त के तिथि से 7 दिनों के अन्दर अवश्य आपूर्ति करना होगा अन्यथा कयादेश रद्द
12. निविदा से संबंधित किसी भी प्रकार के विवाद की स्थिति में न्यायिक क्षेत्र भागलपुर होगा।

डा0 ए0 के0 दास

अधीक्षक

जवाहरलाल नेहरू चिकित्सा महाविद्यालय
अस्पताल, भागलपुर।